



18 August 2015

To: Recipients of M40-A2

From: Jennifer K. Adams, MT(ASCP), MSHA

Subject: Error in Table 9 in CLSI Document M40-A2

This notification is to inform you of an error in Table 9 in CLSI document M40-A2, *Quality Control of Microbiological Transport Systems; Approved Standard—Second Edition*.

Table 9 reference *S. sonnei* ATCC® 12022 as one of the organisms to use in performing QC testing for fecal transport devices.

The organism ATCC® number is incorrect and has been corrected as shown highlighted below:

Table 9. Organisms for Use in Performing QC Testing for Fecal Transport Devices

Species	Strain	Media	Incubation Temperature (°C)	Incubation Atmosphere	Testing Time (hours)
<i>E. coli</i>	ATCC® 25922	TSA with 5% sheep blood agar	35 ± 2	Aerobic	24, 48, or 72
<i>S. typhimurium</i>	ATCC® 14028	TSA with 5% sheep blood agar	35 ± 2	Aerobic	24, 48, or 72
<i>S. sonnei</i>	ATCC® 9290	TSA with 5% sheep blood agar	35 ± 2	Aerobic	24, 48, or 72

Abbreviations: ATCC, American Type Culture Collection; TSA, tryptic soy agar.

If you require any additional clarification regarding this correction, please contact CLSI Customer Service (customerservice@clsi.org).

We appreciate your commitment to CLSI, and regret any inconvenience.