

03-05-2024

Global Tender Notification for procurement of Optical Spectrometer for Optical Coherence Tomography (OCT) application.

GTE Approval No: IISc-GTE-2024-345

Department of Electrical Communication Engineering (ECE), Indian Institute of Science, Bangalore 560012, India

This is a Request for Quote (RFQ) for procurement of Optical Spectrometer for Optical Coherence Tomography (OCT) application as part of a limited tender for the Department of Electrical Communication Engineering (ECE) at Indian Institute of Science, Bangalore.

All interested vendors shall submit a response demonstrating their capabilities to produce the requested equipment to the primary point of contact listed below.

The deadline for submission of proposals is **24th May 2024** by 5:00 PM. Proposals should arrive at the office of **Dr. Balaswamy Velpula, Department of Electrical Communication Engineering, Indian Institute of Science, Bangalore, Karnataka 560012, India.**

The bid should be addressed to: **The Chairman, ECE department, Indian Institute of Science, Bangalore 560012, India.**

Direct all questions concerning the acquisition to **Dr. Balaswamy Velpula at: velpulab@iisc.ac.in**

General Terms and Conditions

- 1. Vendors will be required to submit a technical bid and a commercial bid in two separate sealed envelopes.**
2. The technical bid should contain a compliance table with 4 columns in addition to the ones in the technical requirements table that has been included with this RFQ below. The compliance table should include all the items and be in the same order. The first column should describe your compliance in a "Yes" or "No" response. If "No" the second column should

state, the extent of deviation. The “third” column should state the reasons for the deviation if any. The fourth column can be used to compare your tool with that of your competitors or provide details as requested in the technical requirements table below.

3. In the commercial bid, the price should be inclusive of all discounts.
4. The vendor should have qualified technical service personnel for the equipment based in India (preferably in Bangalore).
5. The lead time for the delivery of the equipment should not be more than 3 months from the date of receipt of our purchase order. It should be clearly mentioned in the technical and commercial proposals.
6. All the quotations must be valid for at least 90 days at the time of submission.
- 7. The quotation should be CIP Bangalore.**
8. List of customers and references: The Bidder should have supplied similar equipment in Central Universities, preferably in centrally Funded Technical Institutes (IITs, IISC, IISER, NIT). Please provide the details and contact information.
9. The Bidder must not be blacklisted/banned/suspended or have a record of any service-related dispute with any organization in India or elsewhere. A declaration to this effect should be provided.
10. In the quote vendors are requested to provide itemized costs for all the associated accessories as options.
11. Please provide itemized quotes for the tool and any attachments/packages. Vendors are encouraged to quote for as many packages as their tool portfolio permits.
12. Please provide information regarding the annual maintenance contract (AMC) beyond the warranty period. Please itemize the year wise AMC as options.

13. Warranty terms, duration and additional warranty options are a must for all the components. Please specify the service plan like whether the local distributor will address the issue or the parent company.
14. Items in addition to those listed in the technical table that you would like to bring to the attention of the committee can be listed at the end of the compliance table.
15. Vendors are encouraged to highlight the advantages of their tools over comparable tools from the competitors.
16. If needed, a meeting for any technical clarifications can be scheduled with the undersigned by sending an email.
17. The Institute reserves the right to accept or reject any bid, or to annul the bidding process and reject all bids, at any time prior to the award of contract without thereby incurring any liability of the affected bidder or bidders.
18. After the award of purchase order, the vendor must provide an Order Acknowledgement within 30 days from the receipt of the Purchase Order.

Technical requirements: Please note that the requirements and options listed below are only guidelines. It does not disbar bids that do not meet the criteria listed. Vendors are requested to quote for equipment that meets the criteria to the best extent possible and list deviations. Deviations are NOT an automatic reason for disqualification. They will be discussed by the technical committee prior to making an informed decision.

Technical Specifications

1. Imaging Depth (in air): 5.1 mm
2. Wavelength Range 1450 - 1690 nm
3. Bandwidth 240 nm
4. Spectral Resolution 0.12 nm
5. Max Line Rate 147 kHz
6. Pixels 2048
7. Interface Camera Link

8. Dimensions 9.5 x 19 x 6 cm
9. Weight 1.4 kg
10. Premier optical design with superior SNR & subpixel resolution for greater image clarity.
11. Superior sensitivity roll-off for 1600 nm: <10 dB @ 3.0 mm imaging depth.
12. 25% greater depth penetration & better SNR for highly scattering samples than 1300 nm.
13. Choice of scan rate, 76 & 147 kHz
14. Proprietary VPH gratings for low polarization dependence & highest efficiency.
15. OEM ready: compact & robust.

Other requirements:

1. To perform installation at the customer site. To provide training to users at customer site.
2. Supplier should agree to provide Performance test reports prior to dispatch of goods.
3. Compatible operating system(s) for the interface software should be specified. Suitable software drivers available should be specified.
4. Please include other options currently available which can be added on in the future.
5. The cost of shipping to IISc should be included.
6. List of acceptance tests for on-site (vendor) inspection and after installation at IISc.
7. A set of basic experiments for performing routine checks of acceptable operation with clear instructions to be provided.
8. The payment terms will be specified in the commercial proposal and is subject to negotiations.
9. Please provide details of the number of trained personnel in India, number in southern region or in Bangalore who can service the instrument.
10. Service credentials: The supplier should have at least five similar installations in India.
11. Customer list with contact details mandatory to prove your credential.
12. Authorisation letter from OEM manufacturer to be included.
13. Vendor must provide complete compliance statement against each technical point.